

WEINSCHTEL ENGINEERING

LAITHEIMSBURG, MARYLAND

Certificate of Calibration

MODEL NO. 1506N SERIAL NO. 23453

FREQ. KMC	ATT DB		1 PORTS		2 FEMALE PORTS	
	1-2	1-3	1 MALE		2 FEMALE C3	
D.C.	5.99	6.00				
3.0	6.0	5.9	1.07		1.05	1.05
6.0	6.2	6.1	1.05		1.03	1.05
10.0	6.4	6.3	1.15		1.06	1.06

Tester B. Koffe

Inspector (Signature)

CALIBRATION STANDARDS AS APPLICABLE ARE TRACEABLE TO THE NATIONAL BUREAU OF STANDARDS