



# Weinschel Engineering

GAITHERSBURG, MARYLAND

## TEST REPORT

Type 89-909-7  
SIN 8346

Job No. \_\_\_\_\_

Customer: \_\_\_\_\_

Date: \_\_\_\_\_

LOT NO.	SERIAL NUMBER	FREQ. <u>DC</u>			FREQ. <u>1.0GHz</u>			FREQ. <u>2.0GHz</u>			FREQ.		
		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.	
			MALE	FEMALE		MALE	FEMALE		MALE	FEMALE		MALE	FEMALE
1	<u>0</u>	<u>.01</u>			<u>.08</u>			<u>.12</u>					
2	<u>10</u>	<u>9.9</u>			<u>10.2</u>			<u>10.1</u>					
3	<u>20</u>	<u>19.96</u>			<u>20.2</u>			<u>20.1</u>					
4	<u>30</u>	<u>29.77</u>			<u>30.0</u>			<u>29.9</u>					
5	<u>40</u>	<u>39.67</u>			<u>39.9</u>			<u>39.7</u>					
6	<u>50</u>	<u>49.56</u>			<u>49.8</u>			<u>49.6</u>					
7	<u>60</u>	<u>59.51</u>			<u>59.5</u>			<u>59.4</u>					
8	<u>70</u>	<u>69.76</u>			<u>69.8</u>			<u>69.6</u>					
9	<u>80</u>	<u>79.4</u>			<u>79.3</u>			<u>79.2</u>					
10	<u>90</u>	<u>90.0</u>			<u>89.6</u>			<u>89.6</u>					
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													

TESTER D. Bell  
 QUALITY CONTROL EB