



Weinschel Engineering

GAITHERSBURG, MARYLAND

TEST REPORT

Type AC 9003-69-31

S/N 4750

Job No. _____

Customer: _____

Date: _____

LOT NO.	SERIAL NUMBER	FREQ. DC				FREQ. 4.0 GHz				FREQ.			
		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.		ATT. DB	V. S. W. R.	
			MALE	FEMALE		MALE	FEMALE		MALE	FEMALE		MALE	FEMALE
1	0	.02			.2								
2	1	.81			.90								
3	2	1.84			1.9								
4	3	2.84			2.9								
5	4	3.85			3.9								
6	5	4.83			4.9								
7	6	5.82			6.0								
8	7	6.79			6.8								
9	8	7.82			7.8								
10	9	8.82			8.8								
11	10	9.82			9.9								
12	20	19.86			20.0								
13	30	29.78			29.9								
14	40	39.70			39.8								
15	50	49.65			49.7								
16	60	59.50			59.5								
17													
18													
19													
20													
21													
22													
23													
24													

TESTER *P. Trivett*
 QUALITY CONTROL *E. Bailey*